



# Certificate of Calibration

CUSTOMER: INDO LIGHTING LTD CERT No: 0348662  
ORDER NO: CUST. REF: NO CUST REF  
MAKE: HAMEG TYPE: HM8115-2  
DESCRIPTION: PROGRAMMABLE POWER METER SERIAL No: 061080009  
AMBIENT TEMPERATURE\*: 20 ± 3 °C HUMIDITY: 55 ± 20 %RH

This is to certify the above instrument has been calibrated in accordance with the relevant specification and at the points tested the results were\*:

Found to meet that specification on receipt [ <input checked="" type="checkbox"/> ]	Found to meet that specification after adjustment/repair [ <input type="checkbox"/> ]	Measurements recorded in absence of relevant specification [ <input type="checkbox"/> ]	Found NOT to meet that specification – Calibration restrictions apply [ <input type="checkbox"/> ]
Pre-Calibration repair performed [ <input type="checkbox"/> ]	Optimising adjustment performed [ <input type="checkbox"/> ]	Calibration performed by subcontractor* [ <input type="checkbox"/> ]	

Absolute Calibration is registered under BS EN ISO 17025:2005 and BS EN ISO 9001:2008

\*For calibration performed by a subcontractor please see the attached certificate for environmental conditions and calibration/measurement details.

In order to comply with the above standards Absolute Calibration has to ensure that all measurements carried out in its laboratories are traceable to national standards.

**Approved Signatory**

DATE: 25/02/13

**Absolute Calibration Limited**

14 Murrills Estate, Portchester, Hampshire, England, PO16 9RD

T: 023 9232 1712 F: 023 9221 0034

W: absolutecal.co.uk E: calit@absolute-cal.co.uk



**Absolute Calibration Limited**

14 Murrills Estate, Portchester, Hampshire, England, PO16 9RD  
 Telephone 023-92321712. Facsimile 023-92210034  
 Service Fax 023-92327100. e-mail calif@absolute-cal.co.uk



Report No.  
**0348662**

**Calibration Laboratory Report**

Laboratory Ambient Temperature  $20 \pm 3$  °C Laboratory Ambient Humidity  $55 \pm 20$  %RH

Cal. Procedure

Description: PROGRAMMEBLE POWER METER

Man'f H/Book

Make: HAMEG

Type: HM8115-2

Classification

Serial No: **061080009**

Cust. Ref:

**(A) B C R**

Customer: **INDO LIGHTING LTD**

Order No:

Code: **IND009**

Test Step No	Function Tested	Nominal Result	Measured Value		Out of Tol	Manufacturers Specification
			As Found	After Adjust		
	<u>DC VOLTAGE RANGE</u>	<u>I/P V</u>	<u>IND V</u>			
	50V	50.0	49.9			
	150V	100	100			
		150	149			
	500V	200	200			$\pm 0.6\%$ + 5 DIGITS
		300	299			
		400	400			
		500	499			
	<u>AC VOLTAGE RANGE</u>	<u>I/P V</u>	<u>IND V</u>			
	50Hz 50V	50.0	50.0			
	150V	100	100			
		150	150			$\pm 0.4\%$ + 5 DIGITS
	500V	200	199			
		300	299			
		400	399			
		500	499			

All measurements are traceable to National Standards

Calibrated By: *R. Nuttall*

Verified By: *[Signature]*

Date: **25/02/13**

HM8115-2.DOC

Issue 1

Page 1 of 4

Test Step No	Function Tested	Nominal Result	Measured Value		Out of Tol	Manufacturers Specification
			As Found	After Adjust		
	<u>DC CURRENT RANGE</u>	<u>I/PA</u>	<u>IND A</u>			
	0.16A	0.100	0.100			
		0.150	0.150			
	1.6A	1.000	0.998			
		1.500	1.499			± 0.6%
	16A	2.00	1.97			+ 5 DIGITS
		4.00	3.98			
		6.00	5.99			
		8.00	7.98			
		10.00	9.99			
	<u>AC CURRENT RANGE</u> 50Hz	<u>I/PA</u>	<u>IND A</u>			
	0.16A	0.100	0.100			
		0.150	0.150			
	1.6A	1.000	0.998			
		1.500	1.499			± 0.4%
	16A	2.00	1.99			+ 5 DIGITS
		4.00	4.00			
		6.00	6.00			
		8.00	8.01			
		10.00	10.02			

Test Step No	Function Tested	Nominal Result	Measured Value		Out of Tol	Manufacturers Specification
			As Found	After Adjust		
	<u>POWER/WATTS</u> 50Hz	<u>I/P W</u>	<u>IND W</u>			
	10V 1A	10	9.93			
	20V 1A	20	19.93			
	30V 1A	30	29.93			
	40V 1A	40	39.91			
	50V 1A	50	49.88			
	100V 1A	100	99.9			
	200V 1A	200	199.7			
	200V 2A	400	400			
	250V 2A	500	499			
	500V 2A	1000	996			
	500V 4A	2000	1999			
	500V 6A	3000	3000			
	500V 8A	4000	4002			
	500V 10A	5000	5005			
						± 0.5% + 10 DIGITS

Earth Bonding and Electrical Safety Test	
Visual Inspection Ck (✓)	Mains Lead Ck (NA)
Earth Continuity 0.152Ω	Load Test <0.05KVA
Insulation Test >300MΩ	Leakage Test <0.30mA

Instruments used in this Calibration

ACL CODING	RECAL DUE
C2551	11/13
N0009	10/13

UNCERTAINTIES OF MEASUREMENT

DC VOLTAGE APPLIED ± 0.04%
AC VOLTAGE APPLIED ± 0.03%
DC CURRENT APPLIED ± 0.07%
AC CURRENT APPLIED ± 0.35%