


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>0452</p> <p>Accredited to ISO/IEC 17025:2005</p>	<p style="text-align: center;">Electronic Test and Calibration Ltd</p> <p style="text-align: center;">Issue No: 022 Issue date: 20 October 2014</p>	
	<p>Caddesdown Industrial Estate Clovelly Road Bideford Devon EX39 3DX</p>	<p>Contact: Steve Campion Tel: +44 (0)1237 423388 Fax: +44 (0)1237 423434 E-Mail: info@etcal.co.uk Website: www.etcal.co.uk</p>

Calibration performed at the above address only

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
DC RESISTANCE	100 $\mu\Omega$ to 10 $M\Omega$ 10 $m\Omega$ to 1.0 Ω 1.0 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2.0 $k\Omega$ 2.0 $k\Omega$ to 20 $k\Omega$ 20 $k\Omega$ to 200 $k\Omega$ 200 $k\Omega$ to 2.0 $M\Omega$ 2.0 $M\Omega$ to 10 $M\Omega$	270 ppm + 0.30 $\mu\Omega$ 120 ppm + 1.0 $\mu\Omega$ 15 ppm + 80 $\mu\Omega$ 12 ppm 11 ppm 12 ppm 11 ppm 28 ppm 30 ppm	Measurement and generation
	10 $M\Omega$ to 2.5 $G\Omega$ 2.5 $G\Omega$ to 250 $G\Omega$ 250 $G\Omega$ to 2.0 $T\Omega$	0.040 % 0.060 % 0.15 %	Measurement only
	10 Ω 100 Ω 1.0 $k\Omega$ 10 $k\Omega$ 100 $k\Omega$ 1.0 $M\Omega$ 10 $M\Omega$ 10 $M\Omega$ to 2.5 $G\Omega$ 2.5 $G\Omega$ to 250 $G\Omega$ 250 $G\Omega$ to 2.0 $T\Omega$	10 ppm 8.0 ppm 8.0 ppm 8.0 ppm 8.0 ppm 20 ppm 20 ppm 0.040 % 0.060 % 0.15 %	Generation only
DC VOLTAGE	0 mV to 200 mV 200 mV to 2.0 V 2.0 V to 20 V 20 V to 200 V 200 V to 1000 V 1.0 kV to 40 kV	13 ppm + 0.50 μV 10 ppm 10 ppm 11 ppm 14 ppm 0.15 %	Sources available up to 30 kV
DC CURRENT	20 μA to 200 μA 200 μA to 200 mA 200 mA to 1.0 A 1.0 A to 10 A 10 A to 100 A	20 ppm 17 ppm 27 ppm 100 ppm 0.15 %	Measurement and generation



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Electronic Test and Calibration Ltd
Issue No: 022 Issue date: 20 October 2014

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
DC CURRENT (cont'd)	10 pA to 2.0 nA 2.0 nA to 200 nA 200 nA to 20 μ A 100 A to 250 A 250 A to 1000 A	0.090 % 0.045 % 0.020 % 0.17 % 0.31 %	Measurement only
AC VOLTAGE	10 pA to 2.0 nA 2.0 nA to 200 nA 200 nA to 20 μ A 100 mHz to 10 Hz 2.5 mV _{rms} to 707 V _{rms} (1000 V _{pk}) 10 Hz to 30 Hz 200 mV to 2.0 V 2.0 V to 20 V 20 V to 200 V 30 (40) Hz to 300 Hz 2.0 mV to 200 mV 200 mV to 2.0 V 2.0 V to 20 V 20 V to 200 V 200 V to 1000 V 300 Hz to 1.0 kHz 2.0 mV to 200 mV 200 mV to 2.0 V 2.0 V to 20 V 20 V to 200 V 200 V to 1000 V 1.0 kHz to 10 kHz 2.0 mV to 200 mV 200 mV to 2.0 V 2.0 V to 20 V 20 V to 200 V 200 V to 1000 V 10 kHz to 30 kHz 2.0 mV to 200 mV 200 mV to 2.0 V 2.0 V to 20 V 20 V to 200 V 200 V to 1000 V 30 kHz to 100 kHz 2.0 mV to 200 mV 200 mV to 2.0 V 2.0 V to 20 V 20 V to 200 V 200 V to 700 V	0.12 % 0.050 % 0.030 % 0.15 % + 5.0 μ V 200 ppm 200 ppm 200 ppm 190 ppm + 5.0 μ V 120 ppm 125 ppm 125 ppm 140 ppm 170 ppm + 5.0 μ V 120 ppm 130 ppm 130 ppm 140 ppm 200 ppm + 5.0 μ V 170 ppm 170 ppm 170 ppm 180 ppm 400 ppm + 5.0 μ V 300 ppm 300 ppm 300 ppm 300 ppm 600 ppm + 5.0 μ V 350 ppm 350 ppm 500 ppm 550 ppm	Generation only



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Electronic Test and Calibration Ltd
Issue No: 022 Issue date: 20 October 2014

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
AC VOLTAGE (cont'd)	<i>100 kHz to 300 kHz</i> 200 mV to 2.0 V 2.0 V to 20 V <i>300 kHz to 1.0 MHz</i> 200 mV to 2.0 V 2.0 V to 20 V <i>40 Hz to 1000 Hz</i> 1.0 kV to 4.0 kV <i>At 50 Hz</i> 1.0 kV to 27 kV	0.10 % 0.10 % 0.50 % 0.50 % 1.0 % 0.30 %	Sources are available up to 6.0 kV at 50 Hz only
AC CURRENT	<i>10 Hz to 30 Hz</i> 10 uA to 200 uA 200 uA to 2.0 mA 2.0 mA to 20 mA 20 mA to 200 mA 200 mA to 1.0 A <i>30 Hz to 1.0 kHz</i> 10 uA to 200 uA 200 uA to 2.0 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 1.0 A 1.0 A to 10 A <i>1.0 kHz to 5.0 kHz</i> 10 uA to 200 uA 200 uA to 2.0 mA 2.0 mA to 20 mA 20 mA to 200 mA 200 mA to 1.0 A 1.0 A to 10 A	200 ppm 200 ppm 200 ppm 200 ppm 250 ppm 200 ppm 200 ppm 200 ppm 200 ppm 250 ppm 310 ppm 300 ppm 300 ppm 300 ppm 300 ppm 400 ppm 500 ppm	
CAPACITANCE	<i>Frequency 100 Hz and 1.0 kHz</i> 10 pF to 100 nF 100 pF to 1.0 μ F 1.0 μ F to 10 μ F 10 μ F to 100 μ F	0.20 % 0.050 % 0.070 % 0.30 %	Measurement and generation
DISSIPATION FACTOR	<i>Frequency 1.0 kHz</i> 0 to 0.01 0.01 to 0.1 0.1 to 0.5 0.5 to 1.0	0.000 26 0.000 27 0.000 54 0.000 85	Dissipation factor ($\tan \delta$) can be reported for capacitance values in the range 1.0 nF to 10 μ F
INDUCTANCE	<i>At 100 Hz and 1.0 kHz</i> 1.0 μ H to 10 μ H 10 μ H to 100 μ H 100 μ H to 10 H	0.90 % 0.30 % 0.15 %	Measurement capability only. The generation of known inductance values may be undertaken over the range of 1.0 mH to 1.0 H but the uncertainties may be increased.



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Electronic Test and Calibration Ltd
Issue No: 022 Issue date: 20 October 2014

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
FREQUENCY MEASUREMENT			Can be expressed as time (1/f) for repetitive wave forms
Specific Values	1.0 MHz and 10 MHz	1 in 10^{10}	
Other Values	1.0 Hz to 1.0 GHz 1.0 GHz to 26.5 GHz	16 in 10^9 16 in 10^9	Generation limited to 20 GHz
TIME INTERVAL	10^{-5} s to 1.0 ns	16 in $10^9 + 1.3$ ns (Full waveform) 17 in $10^9 + 1.3$ ns (Part waveform)	Single Event
HARMONIC AMPLITUDE	30 mV to 300 V	5.8 % of full-scale deflection	Full-scale ranges from 30 mV to 300 V in 3-10-30 sequence. All frequency components must be in the range 10 Hz to 76 kHz.
DISTORTION FACTOR	300 mV to 100 V	3.5 % of full-scale deflection	Fundamental frequency range 20 Hz to 20 kHz; distortion component range up to 100 kHz. Full-scale ranges from 300 mV to 100 V in 3-10-30 sequence.
TEMPERATURE	0 °C to 100 °C	0.25 °C	As a support activity for determination of reference junction temperature when simulating temperature measurements electrically
ELECTRICAL SIMULATION OF TEMPERATURE			
Thermocouple			Measurement and Generation linearity
Type K	-270 °C to 1372 °C	0.050 % + 2.0 μ V	
Type J	-210 °C to 1200 °C	0.050 % + 2.0 μ V	
Type E	-270 °C to 1000 °C	0.050 % + 2.0 μ V	
Type N	-270 °C to 1300 °C	0.050 % + 2.0 μ V	
Type T	-270 °C to 400 °C	0.050 % + 2.0 μ V	
Type S	0 °C to 1768 °C	0.050 % + 2.0 μ V	
Type R	0 °C to 1768 °C	0.050 % + 2.0 μ V	
Type B	0 °C to 1820 °C	0.050 % + 2.0 μ V	
Thermocouple CJC		0.30 °C	
PRT	-200 °C to 850 °C	0.050 % + 10 m Ω	



Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Electronic Test and Calibration Ltd
Issue No: 022 Issue date: 20 October 2014

Calibration performed at main address only



Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Electronic Test and Calibration Ltd
Issue No: 022 **Issue date:** 20 October 2014

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
VOLTAGE REFLECTION COEFFICIENT	<i>5.0 MHz to 1.0 GHz</i> 0.00 to 0.05 0.05 to 0.1 0.1 to 0.2 0.2 to 0.7 0.7 to 1.0 <i>1.0 GHz to 2.0 GHz</i> 0.00 to 0.05 0.05 to 0.1 0.1 to 0.2 0.2 to 0.7 0.7 to 1.0 <i>2.0 GHz to 5.0 GHz</i> 0.00 to 0.05 0.05 to 0.1 0.1 to 0.2 0.2 to 0.7 0.7 to 1.0 <i>5.0 GHz to 10 GHz</i> 0.00 to 0.05 0.05 to 0.1 0.1 to 0.2 0.2 to 0.7 0.7 to 1.0 <i>10 GHz to 15 GHz</i> 0.00 to 0.05 0.05 to 0.1 0.1 to 0.2 0.2 to 0.7 0.7 to 1.0 <i>15 GHz to 18 GHz</i> 0.00 to 0.05 0.05 to 0.1 0.1 to 0.2 0.2 to 0.7 0.7 to 1.0	0.011 0.015 0.028 0.089 0.12 0.017 0.018 0.026 0.068 0.091 0.021 0.034 0.065 0.22 0.32 0.026 0.028 0.038 0.11 0.14 0.033 0.035 0.042 0.093 0.13 0.035 0.038 0.050 0.13 0.18	50 ohm systems only
RF ATTENUATION (Insertion loss method)	0 dB to 80 dB <i>10 kHz to 1.0 GHz</i> 0 dB to 60 dB <i>1.0 GHz to 18 GHz</i> 60 dB to 80 dB <i>1.0 GHz to 18 GHz</i>	0.37 dB 0.39 dB 1.0 dB	50 ohm systems only



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Electronic Test and Calibration Ltd
Issue No: 022 Issue date: 20 October 2014

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
RF ATTENUATION (Substitution method)	0 dB to 80 dB 10 kHz to 1.0 GHz	0.32 dB	50 ohm systems only
	80 dB to 120 dB 10 kHz to 1.0 GHz	2.5 dB	
	0 dB to 25 dB 10 MHz to 50 MHz	0.060 dB	
	50 MHz to 1.0 GHz	0.030 dB	
	1.0 GHz to 5.0 GHz	0.040 dB	
	5.0 GHz to 10 GHz	0.050 dB	
	10 GHz to 15 GHz	0.060 dB	
	15 GHz to 18 GHz	0.070 dB	
	25 dB to 60 dB 10 MHz to 50 MHz	0.040 dB	
	50 MHz to 1.0 GHz	0.040 dB	
FREQUENCY MODULATION	1.0 GHz to 5.0 GHz	0.040 dB	Carrier frequency range 250 kHz to 10 MHz; modulation frequency range 20 Hz to 10 kHz
	5.0 GHz to 10 GHz	0.060 dB	
	10 GHz to 15 GHz	0.080 dB	
	15 GHz to 18 GHz	0.10 dB	
	60 dB To 70 dB 10MHz to 50MHz	0.24 dB	
	50MHz to 1.0 GHz	0.24 dB	
	1.0 GHz to 5.0 GHz	0.24 dB	
	5.0 GHz to 10 GHz	0.24 dB	
	10 GHz to 15 GHz	0.24 dB	
	15 GHz to 18 GHz	0.25 dB	
FREQUENCY MODULATION	0 Hz to 40 kHz	2.6 % + 2.0 Hz	Carrier frequency range 10 MHz to 1.3 GHz; modulation frequency range 20 Hz to 200 kHz
	0 Hz to 400 kHz	1.6 % + 16 Hz	
	0 Hz to 300 kHz	6.0 % + 200 Hz	



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Electronic Test and Calibration Ltd
Issue No: 022 Issue date: 20 October 2014

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
AMPLITUDE MODULATION	0 to 0.05 0.05 to 0.3 0.3 to 0.5 0.5 to 0.9 0 to 0.05 0.05 to 0.3 0.3 to 0.5 0.5 to 0.9 0 to 0.05 0.05 to 0.3 0.3 to 0.5 0.5 to 0.9	4.3 % + 0.00020 3.1 % + 0.00020 2.8 % + 0.0020 2.6 % + 0.0020 3.7 % + 0.00020 3.7 % + 0.00020 1.8 % + 0.0020 1.6 % + 0.0020 5.6 % + 0.0020 4.8 % + 0.0020 4.6 % + 0.0050 4.6 % + 0.0050	Ranges and uncertainties are shown in terms of modulation index Carrier frequency range 150 kHz to 10 MHz; modulation frequency range 20 Hz to 10 kHz. Carrier frequency range 10 MHz to 1.3 GHz; modulation frequency range 20 Hz to 100 kHz. Carrier frequency range 1.0 GHz to 18 GHz; modulation frequency range 200 Hz to 100 kHz.
PHASE MODULATION	0 radians to 400 radians	3.7 % + 0.0020 radians	Carrier frequency range 10 MHz to 1.3 GHz; modulation frequency range 20 Hz to 20 kHz
RF INTERMODULATION PRODUCTS	0 dB to - 80 dB <i>10 kHz to 110 MHz</i> <i>110 MHz to 18 GHz</i>	0.94 dB 1.9 dB	
PULSE RISETIME AND FALLTIME	1.0 ns to 1.0 s	2.9 %	For the calibration of Waveform Generators
PULSE WIDTH	1.0 ns to 1.0 s	1.2 %	
VOLTAGE AMPLITUDE	0.1 kV to 6.6 kV	3.0 %	
ELECTROSTATIC VOLTAGE	0.1 kV to 30 kV	1.0 %	Fieldmeters for measuring charged surfaces
HIGH IMPEDANCE CONTACT VOLTAGE	0.1 kV to 30 kV	0.70 %	Electrostatic voltmeter and other high resistance voltmeters for measuring charged surfaces
ELECTROSTATIC DISCHARGE GENERATORS			
EN 61000-4-2:1995			
PULSE AMPLITUDE	0.1 kV to 30 kV	1.0 %	Contact voltmeter
PULSE AMPLITUDE	0.1 A to 30 A	2.8 %	ESD Target 1 st peak
RISE TIME	100 ps to 1.0 s	3.7 %	ESD Target
FALL TIME AMPLITUDE	0.1 A to 30 A	3.8 %	ESD Target



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Electronic Test and Calibration Ltd
Issue No: 022 Issue date: 20 October 2014

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
ELECTROSTATIC DISCHARGE GENERATORS (cont'd)			
EN 61000-4-2:1995			
FALL TIME	100 ps to 1.0 s	3.7 %	ESD Target
EN 61000-4-2:2009			
PULSE AMPLITUDE	0.1 kV to 30 kV	1.0 %	Contact voltmeter
PULSE AMPLITUDE	0.1 A to 50 A	5.9 %	ESD Target 1 st peak
RISE TIME	200 ps to 1.0 s	100 ps	ESD Target
FALL TIME AMPLITUDE	0.1 A to 50 A	5.9 %	ESD Target
FALL TIME	200 ps to 1.0 s	100 ps	ESD Target
Voltage amplitude	0.5 kV to 4.4 kV	2.9 %	
BURST TRANSIENT GENERATOR CHARACTERISTICS			
Risetime and falltime	1 ns to 1.0 s	1.7 %	For the calibration of Electrical Fast Transient generators and CDN's to 61000-4-4
Pulse width	0.6 ns to 1.0 ms	1.2 %	
Repetition rate	1 ns to 100 s	0.03 %	
Frequency	1 Hz to 1 GHz	0.3 %	
Voltage amplitude	0.25 kV to 6.6 kV	3.0 %	
SURGE PULSE CHARACTERISTICS			
Risetime and falltime	0.1 μ s to 1.0 s	2.3 %	For the calibration of Surge generators to 61000-4-5
Pulse width	0.6 μ s to 1.0 ms	1.5 %	
Repetition rate	1.0 s to 100 s	0.03 %	
Phase angle	0° to 360°	1.3°	
Voltage amplitude	0.25 kV to 6.6 kV	2.6 %	
Current amplitude	0.2 kA to 3.3 kA	2.9 %	
RF IMPEDANCE	5.0 Ω to 60 Ω 9.0 kHz to 30 MHz	4.6 %	For impedance calibration of line impedance stabilisation networks (LISNs)
IMPULSE GENERATOR MEASUREMENTS	50 dB μ V to 110 dB μ V 9.0 kHz to 30 MHz 30 MHz to 300 MHz 300 MHz to 1.0 GHz	0.7 dB 1.0 dB 1.5 dB	



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Electronic Test and Calibration Ltd
Issue No: 022 Issue date: 20 October 2014

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
IMPULSE GENERATION			
Absolute and relative amplitude	50 dB μ V to 110 dB μ V 9.0 kHz to 150 kHz	0.25 dB	
Absolute amplitude	50 dB μ V to 110 dB μ V 150 kHz to 30 MHz 30 MHz to 300 MHz 300 MHz to 1.0 GHz	0.80 dB 1.1 dB 1.6 dB	
Relative amplitude	50 dB μ V to 110 dB μ V 150 kHz to 30 MHz 30 MHz to 300 MHz 300 MHz to 1.0 GHz	0.50 dB 0.60 dB 0.90 dB	
Voltage Dips, Short Interruptions Voltage Variations Generators			
Dip RMS Voltage	1 V to 400 V	2.3 %	
Voltage Variations	1 V to 400 V	3.5 %	
Interruptions Overshoot Voltage	25 % to 100 %	3.5 %	
Transition rise and fall time	1.0 μ s to 1.0 s	3.0 %	
Phase Angle	0° to 360°	11°	
Damped Oscillatory Generator			
Voltage	1.0 V to 6.6 kV	2.7 %	
Ringwave Current	1.0 A to 3.3 kA	2.9 %	
Oscillatory Wave Current	1.0 A to 150 A	4.4 %	
Rise time	1.0 ns to 1.0 s	3.0 %	
Frequency	1 Hz to 1 GHz	2.3 %	
Repetition Rate	1.0 μ s to 1.0 s	2.3 %	
Duration	10 ns to 10 s	2.3 %	
Period	1.0 ns to 1.0 s	1.0 %	
Phase	0° to 360°	11°	
ANTENNA MEASUREMENTS			
Monopole Antenna Antenna Factor	20 Hz to 30 MHz 30 MHz to 100 MHz	1.4 dB/m 1.6 dB/m	Equivalent capacitance method
Antenna Factor & Apparent Gain			Best capability using the three antenna method or by comparison with similar antennas using the standard antenna method
Biconical and Broad Band Dipoles	20 MHz to 300 MHz 300 MHz to 1.0 GHz	1.2 dB 1.2 dB	Measurement distance 10 m, 3.0 m and 1.0 m
Log Periodic	80 MHz to 18.0 GHz	1.2 dB	Measurement distances 3.0 m and 1.0 m; calculated results for 10 m and for Free Space



Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Electronic Test and Calibration Ltd
Issue No: 022 Issue date: 20 October 2014

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
ANTENNA MEASUREMENTS (cont)			
Bilog and hybrid antennas	20 MHz to 18.0 GHz	1.4 dB	Measurement distances 3.0 m and 1.0 m; calculated results for 10 m and for Free Space
Horn Antennas	200 MHz to 1.0 GHz 1.0 GHz to 18.0 GHz	1.5 dB 1.2 dB	Horn measurement at 3.0 m and 1.0 m
END			